

MC-21-C电阻箱出厂检测报告					
生产厂家:泸州泸电科技有限公司			官网: http://ldkj.ywtya.com		
测试机身编号:	MC-21-C 20231124223		测试环境:	温度:15° 湿度:80%	
测试日期:	2023年12月8日		测试人:	江明友	
测试设备:	三杰 SJ2515P 三量RT640 (M以上)		全档0位阻值:	0.039Ω (左侧红黑接线柱)	
说明: 所有单位均为Ω 实测值为0位阻值归零后测试设备实显值,已除掉0位阻值			测试结果: 合格		
档位(标称值)	实测值	标称误差	合格范围	合格状态	备注
1	1.022	5%	0.95~1.05	PASS	
2	2.021	5%	1.9~2.1	PASS	
3	3.007	5%	2.85~3.15	PASS	
4	4.025	5%	3.8~4.2	PASS	
5	5.023	5%	4.75~5.25	PASS	
6	6.003	5%	5.7~6.3	PASS	
7	7.007	5%	6.65~7.35	PASS	
8	8.049	5%	7.6~8.4	PASS	
9	9.048	5%	8.55~9.45	PASS	
10	10.019	2%	9.8~10.2	PASS	
20	20.041	2%	19.6~20.4	PASS	
30	30.034	2%	29.4~30.6	PASS	
40	40.072	2%	39.2~40.8	PASS	
50	50.041	2%	49~51	PASS	
60	60.019	2%	58.8~61.2	PASS	
70	70.101	2%	68.6~71.4	PASS	
80	80.049	2%	78.4~81.6	PASS	
90	90.046	2%	88.2~91.8	PASS	
100	99.99	2%	98~102	PASS	
200	199.95	2%	196~204	PASS	
300	300.08	2%	294~306	PASS	
400	399.82	2%	392~408	PASS	
500	499.78	2%	490~510	PASS	
600	600.62	2%	588~612	PASS	
700	700.46	2%	686~714	PASS	
800	800.23	2%	784~816	PASS	
900	898.91	2%	882~918	PASS	
1K	0.999K	2%	0.98K~1.02K	PASS	
2K	2.001K	2%	1.96K~2.04K	PASS	
3K	3.001K	2%	2.94K~3.06K	PASS	
4K	4.001K	2%	3.92K~4.08K	PASS	
5K	4.999K	2%	4.9K~5.1K	PASS	
6K	5.997K	2%	5.88K~6.12K	PASS	
7K	7.001K	2%	6.86K~7.14K	PASS	
8K	7.998K	2%	7.84K~8.16K	PASS	
9K	9.005K	2%	8.82K~9.18K	PASS	
10K	9.998K	2%	9.8K~10.2K	PASS	
20K	19.997K	2%	19.6K~20.4K	PASS	
30K	29.976K	2%	29.4K~30.6K	PASS	
40K	40.155K	2%	39.2K~40.8K	PASS	
50K	50.095K	2%	49K~51K	PASS	
60K	60.046K	2%	58.8K~61.2K	PASS	
70K	70.157K	2%	68.6K~71.4K	PASS	
80K	80.066K	2%	78.4K~81.6K	PASS	
90K	90.209K	2%	88.2K~91.8K	PASS	
100K	100.15K	2%	98K~102K	PASS	
200K	200.79K	2%	196K~204K	PASS	
300K	300.01K	2%	294K~306K	PASS	
400K	400.45K	2%	392K~408K	PASS	
500K	500.21K	2%	490K~510K	PASS	
600K	601.96K	2%	588K~612K	PASS	
700K	701.13K	2%	686K~714K	PASS	
800K	801.32K	2%	784K~816K	PASS	
900K	903.37K	2%	882K~918K	PASS	
1M	0.96M	10%	0.9M~1.1M	PASS	测试电压:500V
2M	1.92M	10%	1.8M~2.2M	PASS	测试电压:500V
3M	2.91M	10%	2.7M~3.3M	PASS	测试电压:500V
4M	3.87M	10%	3.6M~4.4M	PASS	测试电压:500V
5M	4.85M	10%	4.5M~5.5M	PASS	测试电压:500V
6M	5.82M	10%	5.4M~6.6M	PASS	测试电压:500V
7M	6.79M	10%	6.3M~7.7M	PASS	测试电压:500V
8M	7.77M	10%	7.2M~8.8M	PASS	测试电压:500V
9M	8.75M	10%	8.1M~9.9M	PASS	测试电压:500V
10M	9.9M	10%	9M~11M	PASS	测试电压:500V
20M	19.9M	10%	18M~22M	PASS	测试电压:500V
30M	29.8M	10%	27M~33M	PASS	测试电压:500V
40M	39.8M	10%	36M~44M	PASS	测试电压:500V
50M	49.8M	10%	45M~55M	PASS	测试电压:500V
60M	59.9M	10%	54M~66M	PASS	测试电压:500V
70M	69.9M	10%	63M~77M	PASS	测试电压:500V
80M	79.8M	10%	72M~88M	PASS	测试电压:500V
90M	90.1M	10%	81M~99M	PASS	测试电压:500V
100M	101M	10%	90M~110M	PASS	测试电压:1000V
200M	202M	10%	180M~220M	PASS	测试电压:1000V
300M	302M	10%	270M~330M	PASS	测试电压:1000V
400M	404M	10%	360M~440M	PASS	测试电压:1000V
500M	506M	10%	450M~550M	PASS	测试电压:1000V
600M	608M	10%	540M~660M	PASS	测试电压:1000V
700M	711M	10%	630M~770M	PASS	测试电压:1000V
800M	814M	10%	720M~880M	PASS	测试电压:1000V
900M	916M	10%	810M~990M	PASS	测试电压:1000V
1G	1.06G	10%	0.9G~1.1G	PASS	测试电压:2500V
2G	2.12G	10%	1.8G~2.2G	PASS	测试电压:2500V
3G	3.18G	10%	2.7G~3.3G	PASS	测试电压:2500V
4G	4.21G	10%	3.6G~4.4G	PASS	测试电压:2500V
5G	5.22G	10%	4.5G~5.5G	PASS	测试电压:2500V
6G	6.29G	10%	5.4G~6.6G	PASS	测试电压:2500V
7G	7.38G	10%	6.3G~7.7G	PASS	测试电压:2500V
8G	8.48G	10%	7.2G~8.8G	PASS	测试电压:2500V
9G	9.4G	10%	8.1G~9.9G	PASS	测试电压:2500V