

MC-21-D电阻箱出厂检测报告					
生产厂家:泸州泸电科技有限公司			官网: http://ldkj.ywtva.com		
测试机身编号: MC-21-D 20231124322		测试环境:		温度:15° 湿度:80%	
测试日期: 2023年12月8日		测试人:		江明友	
测试设备: 三杰 SJ2515P 三量RT640 (M以上)		全档0位阻值:		0.036Ω (左侧红黑接线柱)	
说明: 所有单位均为Ω 实测值为0位阻值归零后测试设备实显值,已除掉0位阻值		测试结果: 合格			
档位(标称值)	实测值	标称误差	合格范围	合格状态	备注
1	1.021	5%	0.95~1.05	PASS	
2	2.015	5%	1.9~2.1	PASS	
3	3.021	5%	2.85~3.15	PASS	
4	4.028	5%	3.8~4.2	PASS	
5	5.009	5%	4.75~5.25	PASS	
6	5.999	5%	5.7~6.3	PASS	
7	7.017	5%	6.65~7.35	PASS	
8	8.055	5%	7.6~8.4	PASS	
9	9.022	5%	8.55~9.45	PASS	
10	10.029	2%	9.8~10.2	PASS	
20	20.053	2%	19.6~20.4	PASS	
30	30.039	2%	29.4~30.6	PASS	
40	40.075	2%	39.2~40.8	PASS	
50	50.037	2%	49~51	PASS	
60	59.991	2%	58.8~61.2	PASS	
70	70.099	2%	68.6~71.4	PASS	
80	80.098	2%	78.4~81.6	PASS	
90	90.069	2%	88.2~91.8	PASS	
100	99.96	2%	98~102	PASS	
200	199.96	2%	196~204	PASS	
300	300.12	2%	294~306	PASS	
400	400.01	2%	392~408	PASS	
500	499.69	2%	490~510	PASS	
600	599.53	2%	588~612	PASS	
700	700.75	2%	686~714	PASS	
800	799.95	2%	784~816	PASS	
900	900.14	2%	882~918	PASS	
1K	0.999K	2%	0.98K~1.02K	PASS	
2K	2.001K	2%	1.96K~2.04K	PASS	
3K	2.999K	2%	2.94K~3.06K	PASS	
4K	4.001K	2%	3.92K~4.08K	PASS	
5K	5.001K	2%	4.9K~5.1K	PASS	
6K	5.996K	2%	5.88K~6.12K	PASS	
7K	7.002K	2%	6.86K~7.14K	PASS	
8K	7.995K	2%	7.84K~8.16K	PASS	
9K	8.993K	2%	8.82K~9.18K	PASS	
10K	10.005K	2%	9.8K~10.2K	PASS	
20K	19.979K	2%	19.6K~20.4K	PASS	
30K	29.995K	2%	29.4K~30.6K	PASS	
40K	40.067K	2%	39.2K~40.8K	PASS	
50K	50.138K	2%	49K~51K	PASS	
60K	60.071K	2%	58.8K~61.2K	PASS	
70K	70.116K	2%	68.6K~71.4K	PASS	
80K	80.135K	2%	78.4K~81.6K	PASS	
90K	90.236K	2%	88.2K~91.8K	PASS	
100K	100.33K	2%	98K~102K	PASS	
200K	200.49K	2%	196K~204K	PASS	
300K	300.04K	2%	294K~306K	PASS	
400K	400.91K	2%	392K~408K	PASS	
500K	500.04K	2%	490K~510K	PASS	
600K	601.87K	2%	588K~612K	PASS	
700K	699.14K	2%	686K~714K	PASS	
800K	800.23K	2%	784K~816K	PASS	
900K	903.85K	2%	882K~918K	PASS	
1M	0.96M	10%	0.9M~1.1M	PASS	测试电压: 500V
2M	1.92M	10%	1.8M~2.2M	PASS	测试电压: 500V
3M	2.89M	10%	2.7M~3.3M	PASS	测试电压: 500V
4M	3.86M	10%	3.6M~4.4M	PASS	测试电压: 500V
5M	4.84M	10%	4.5M~5.5M	PASS	测试电压: 500V
6M	5.81M	10%	5.4M~6.6M	PASS	测试电压: 500V
7M	6.78M	10%	6.3M~7.7M	PASS	测试电压: 500V
8M	7.76M	10%	7.2M~8.8M	PASS	测试电压: 500V
9M	8.73M	10%	8.1M~9.9M	PASS	测试电压: 500V
10M	9.9M	10%	9M~11M	PASS	测试电压: 500V
20M	19.8M	10%	18M~22M	PASS	测试电压: 500V
30M	29.8M	10%	27M~33M	PASS	测试电压: 500V
40M	39.8M	10%	36M~44M	PASS	测试电压: 500V
50M	49.9M	10%	45M~55M	PASS	测试电压: 500V
60M	60M	10%	54M~66M	PASS	测试电压: 500V
70M	69.9M	10%	63M~77M	PASS	测试电压: 500V
80M	79.9M	10%	72M~88M	PASS	测试电压: 500V
90M	90.1M	10%	81M~99M	PASS	测试电压: 500V
100M	101M	10%	90M~110M	PASS	测试电压: 1000V
200M	202M	10%	180M~220M	PASS	测试电压: 1000V
300M	302M	10%	270M~330M	PASS	测试电压: 1000V
400M	404M	10%	360M~440M	PASS	测试电压: 1000V
500M	505M	10%	450M~550M	PASS	测试电压: 1000V
600M	607M	10%	540M~660M	PASS	测试电压: 1000V
700M	709M	10%	630M~770M	PASS	测试电压: 1000V
800M	812M	10%	720M~880M	PASS	测试电压: 1000V
900M	913M	10%	810M~990M	PASS	测试电压: 1000V
1G	1.07G	10%	0.9G~1.1G	PASS	测试电压: 2500V
2G	2.11G	10%	1.8G~2.2G	PASS	测试电压: 2500V
3G	3.12G	10%	2.7G~3.3G	PASS	测试电压: 2500V
4G	4.18G	10%	3.6G~4.4G	PASS	测试电压: 2500V
5G	5.26G	10%	4.5G~5.5G	PASS	测试电压: 2500V
6G	6.35G	10%	5.4G~6.6G	PASS	测试电压: 2500V
7G	7.44G	10%	6.3G~7.7G	PASS	测试电压: 2500V
8G	8.53G	10%	7.2G~8.8G	PASS	测试电压: 2500V
9G	9.5G	10%	8.1G~9.9G	PASS	测试电压: 2500V
10G	10.7G	10%	9G~11G	PASS	测试电压: 2500V
20G	21.6G	10%	18G~22G	PASS	测试电压: 2500V
30G	32.8G	10%	27G~33G	PASS	测试电压: 2500V
40G	42.3G	10%	36G~44G	PASS	测试电压: 2500V
50G	53.5G	10%	45G~55G	PASS	测试电压: 2500V
60G	65.1G	10%	54G~66G	PASS	测试电压: 2500V
70G	76.5G	10%	63G~77G	PASS	测试电压: 2500V
80G	87.3G	10%	72G~88G	PASS	测试电压: 2500V
90G	99G	10%	81G~99G	PASS	测试电压: 2500V