

MC-21-J电阻箱出厂检测报告					
生产厂家:泸州泸电科技有限公司			官网: http://ldkj.ywtya.com		
测试机身编号:	MC-21-J 20231124423	测试环境:	温度:15° 湿度:80%		
测试日期:	2023年12月8日	测试人:	江明友		
测试设备:	三杰 SJ2515P	全档0位阻值:	0.198Ω		
说明: 所有单位均为Ω 实测值为0位阻值归零后测试设备实显值,已除掉0位阻值			测试结果: 合格		
档位(标称值)	实测值	标称误差	合格范围	合格状态	备注
1	1.009	5%	0.95~1.05	PASS	
2	2.018	5%	1.9~2.1	PASS	
3	3.012	5%	2.85~3.15	PASS	
4	4.024	5%	3.8~4.2	PASS	
5	4.998	5%	4.75~5.25	PASS	
6	6.003	5%	5.7~6.3	PASS	
7	6.993	5%	6.65~7.35	PASS	
8	8.02	5%	7.6~8.4	PASS	
9	9.035	5%	8.55~9.45	PASS	
10	10.027	5%	9.5~10.5	PASS	
20	20.008	5%	19~21	PASS	
30	30.043	5%	28.5~31.5	PASS	
40	40.062	5%	38~42	PASS	
50	50.039	5%	47.5~52.5	PASS	
60	59.952	5%	57~63	PASS	
70	70.066	5%	66.5~73.5	PASS	
80	80.08	5%	76~84	PASS	
90	90.038	5%	85.5~94.5	PASS	
100	99.97	5%	95~105	PASS	
200	199.93	5%	190~210	PASS	
300	300.24	5%	285~315	PASS	
400	400.09	5%	380~420	PASS	
500	499.79	5%	475~525	PASS	
600	599.92	5%	570~630	PASS	
700	699.72	5%	665~735	PASS	
800	800.06	5%	760~840	PASS	
900	900.23	5%	855~945	PASS	
1K	0.999K	5%	0.95K~1.05K	PASS	
2K	2.001K	5%	1.9K~2.1K	PASS	
3K	2.999K	5%	2.85K~3.15K	PASS	
4K	3.995K	5%	3.8K~4.2K	PASS	
5K	5.001K	5%	4.75K~5.25K	PASS	
6K	6.002K	5%	5.7K~6.3K	PASS	
7K	6.998K	5%	6.65K~7.35K	PASS	
8K	7.991K	5%	7.6K~8.4K	PASS	
9K	8.988K	5%	8.55K~9.45K	PASS	
10K	10.008K	5%	9.5K~10.5K	PASS	
20K	19.979K	5%	19K~21K	PASS	
30K	30.011K	5%	28.5K~31.5K	PASS	
40K	40.078K	5%	38K~42K	PASS	
50K	50.132K	5%	47.5K~52.5K	PASS	
60K	59.998K	5%	57K~63K	PASS	
70K	70.023K	5%	66.5K~73.5K	PASS	
80K	80.075K	5%	76K~84K	PASS	
90K	90.109K	5%	85.5K~94.5K	PASS	
100K	100.23K	5%	95K~105K	PASS	
200K	200.89K	5%	190K~210K	PASS	
300K	300.29K	5%	285K~315K	PASS	
400K	400.38K	5%	380K~420K	PASS	
500K	500.96K	5%	475K~525K	PASS	
600K	603.61K	5%	570K~630K	PASS	
700K	700.80K	5%	665K~735K	PASS	
800K	801.37K	5%	760K~840K	PASS	
900K	903.34K	5%	855K~945K	PASS	